

Atos

DECLARATION OF CONFORMITY FOR THE PRODUCT(S)

Provox® Luna® ShowerAid™

Basic UDI: 7331791-ADH-A-000-0007-UV

This Declaration of Conformity is issued under the sole responsibility of Atos Medical AB. Atos Medical AB declares that the devices covered by the present declaration are in conformity with the Regulation (EU) 2017/745 of the European Parliament and of the Council of 5 April 2017 on medical devices and the requirements specified herein are fulfilled and, clause 6.6 of Schedule 3 to the Australian Therapeutic Goods (Medical Devices) Regulations 2002 including amendments in effect at the issuance date.

Intended use/purpose:

The Provox Luna ShowerAid is used with the Provox Luna Adhesive while taking a shower to avoid water from entering the stoma. Single patient use.

Hörby, Sweden, date as stated on last page



.....
Martin Richardson, Senior Vice President Operations & Quality
on behalf of the CEO of Atos Medical AB.

Manufacturer: **Atos Medical AB**
Kraftgatan 8, SE-242 35 Hörby
Sweden

Telephone: +46 (0)415 198 00
Email: Info@atosmedical.com
Web: www.atosmedical.com

SRN number: **SE-MF-000000725**

Competent Authority **Medical Products Agency**
Sweden

DECLARATION OF CONFORMITY

FOR THE PRODUCT(S)

7331791-ADH-A-000-0007-UV

REF	Device name	Class*	GMDN code
8016	Provox Luna ShowerAid	I	62047

*Product Classification according to Annex VIII EU Medical Device Regulation 2017/745

In compliance with Therapeutic Goods (Medical Devices) Regulations 2002

Each kind of medical device to which the system has been applied complies with the applicable provisions of the essential principles, the classification rules, and the full quality assurance procedures, at each stage, from the design of the device until its final inspection before being supplied.

In compliance with UK Medical Devices Regulations 2002 as amended

Each kind of medical device to which the system has been applied fulfils the applicable obligations imposed by Annex VII and the device meets the provisions of the applicable section of the Medical Devices Regulations.

Authorized representative/UK Responsible Person:

Atos Medical UK Ltd Tottle Road
Cartwright House
Nottingham
Nottinghamshire NG2 1RT
England, United Kingdom

Common Specification(s) as per Article 9, and other Union Legislation(s)

- No relevant Common Specifications as defined in Article 9 to list.
- No relevant Union Legislations to list
- No European Representative

Document Approvals
Approved Date: 2023-12-12

Approval Task Verdict: Approve	SEHRBHNU Ulrika Svensson, Quality Assurance Specialist (ulrika.svensson-atosmedical@coloplast.com) Issuer 11-Dec-2023 18:37:51 GMT+0000
Approval Task Verdict: Approve	MARRIC Martin Richardson, VP, International Manufacturing, Sweden, Germany & China (martin.richardson-atosmedical@coloplast.com) Management 11-Dec-2023 19:51:50 GMT+0000
Approval Task Verdict: Approve	KARNIL Karolina Nilsson, Head of Regulatory Affairs (karolina.nilsson-atosmedical@coloplast.com) Regulatory 12-Dec-2023 09:40:01 GMT+0000